PRODUCT INFORMATION SHEET

Almond and Sesame on Polystyrol-Surface (for Allergen Swab Test)

PT originated REFERENCE MATERIAL (RM)

Description of the RM

Name / Material of related PT	*DLA ptASW3 (2023) Allergen Swab Test III: Almond, Hazelnut, Molluscs and Sesame – Plates A and B*
Identifier / LOT	*RM 2023-ASW3 AB*
Expiry date	*October 2027*
Matrix	*dried suspension, see table*
Physical description	*coated test areas (polystyrol surface)*
Packaging / weight	*2 petri dishes / Sealed foil pouche*
Storage conditions	*room temperature, long term cooled 2 - 10°C (dry and dark)*
Name and address manufacturer	***DLA - Proficiency Tests GmbH*** *Hauptstr. 80, 23845 Oering, Germany* *Tel. ++49-(0)4532-9183358, Fax. ++49(0)4102-9944976* *eMail. pt@dla-lvu.de, website. www.dla-lvu.de*

Composition of Material

Added amounts of allergenic ingredients, positive in brackets in µg/test surface (approx. 30 cm^2) quantitative ranges given as food item **

Spiked Ingredients on Surface *	Surface A	Surface B
Almond: roasted almonds, ground (Protein 21,1%)	**positive (40-65)**	**negative**
Sesame: Seeds white, dried (Protein 24,5%)	**negative**	**positive (55-95)**

* Protein contents according to laboratory analysis (total nitrogen, Kjeldahl factor F=5,18 for almond protein, F=5,30 for hazelnut protein and F=6,25 for molluscs protein and sesame protein)
**Allergen contents of „food item“ as indicated in the column of ingredients according gravimetric mixing (±) 25%

Reference Values

All reference values were derived from the related proficiency test (PT) round (for more details see final report).

Reference Values qualitative (Consensus Values*)

Methods: Swab Test + ELISA and Lateral Flow / PCR	Surface A	Surface B
Almond ELISA and Lateral Flow	**positive**	**negative**
Almond PCR (informative, n=1)	**positive**	**negative**
Sesame ELISA and Lateral Flow	**negative**	**positive**
Sesame PCR	**negative**	**positive**

* Consensus values ≥ 75% positive or negative results of PT-participants

Intended Use

- for laboratory use only, no food!
- validation of analytical methods
- control sample for quality control charts
- test material for inter-laboratory comparisons
- no medical use

Storage Instructions and Handling Information

The material should be stored as stated on page 1. Before use of the material should be warmed up to room temperature. The content should be mixed thoroughly prior to subsampling (see minimum sample size above). The material should be weighed immediately after opening of the foil pouche/bottle and closed again afterwards in order to avoid uptake of moisture. Frequent freezing and thawing should be avoided as this can lead to changes in the properties of the material. The expiry date of this RM is based on the current knowledge and applies only for proper storage conditions in the originally closed foil pouche/bottle.

Safety Precautions

The usual laboratory safety precautions should be applied when working with the RM considering the type of ingredients, parameters and its concentration.

Data Source

The material was analyzed by 10 laboratories within the related proficiency test (PT). Different analytical methods have been used.

Various

The user must determine the suitability of this RM for its particular purpose. DLA makes no warranty of any kind, expressed or implied, other than its products meet all quality control standards set by DLA. We do not guarantee this product can be used for a special application.

Datum / Date: 5 November 2024

Dr. Matthias Besler-Scharf (EP-Koordinator / PT-Coordinator)
- gezeichnet / signed M. Besler-Scharf

DLA Proficiency Tests GmbH
Hauptstr. 80, 23845 Oering, Germany